

## YSI Photometer Test Kits

Alkalinity, Total (Alkaphot)	0-500 (CaCO <sub>3</sub> )
Alkalinity-M (Alkaphot M)	0-500 (CaCO <sub>3</sub> )
Alkalinity-P (Alkaphot P)	0-500 (CaCO <sub>3</sub> )
Aluminum	0-0.5
Ammonia	0-1.0 (N)
Bromine	0-10.0
Calcium Hardness (Calcicol)	0-500 (CaCO <sub>3</sub> )
Chloride (Chloridol)	0-50,000 (NaCl)
Chlorine DPD 1 (Free Chlorine)	0-5.0
Chlorine DPD 2 (Monochloramine and Dichloramine - requires YPM031)	0-5.0
Chlorine DPD 1&3 (Free, Combined and Total Chlorine)	0-5.0
Chlorine DPD 4 (Total Chlorine)	0-5.0
Chlorine HR (High Range)	0-250
Chlorine Dioxide (DPD)	0-4.0
Chlorine Dioxide LR (Low Range)	0-2.5
Chlorine Dioxide HR (High Range)	2.5-20
Chromium VI (Hexavalent)	0-1.0
Copper (Coppercol) (free,combined, total)	0-5.0
Copper (free)	0.03-5.0
Color (includes turbidity)	10-500
Cyanuric Acid	0-200
Dissolved Oxygen (0.0-0.8 vials)	0.02-0.08
Dissolved Oxygen (0.0-2.0 vials)	0.05-2.0
Fluoride	0-1.5
Hardness (Hardicol)	0-500 (CaCO <sub>3</sub> )
Hydrazine	0-0.5
Hydrogen Peroxide LR	0-2
Hydrogen Peroxide HR	0-100
Iron LR	0-1.0
Iron MR	0-5.0
Iron HR	0-10
Magnesium (Magnecol)	0-100
Manganese	0-0.03
Molybdate LR	0-20
Molybdate HR	0-100
Nickel	0-10
Nitrate	0-20 (N)
Nitrite (N)	0-0.5 (N)
Nitrite (NaNO <sub>2</sub> )	0-1500 (NaNO <sub>2</sub> )
Organophosphonate (OP)	0-20 (PO <sub>4</sub> )
Ozone	0-2.0
pH (phenol red)	6.8-8.4
PHMB	2-100

Phenol	0-5.0
Phosphate LR	0-4.0
Phosphate HR	0-100
Potassium	0-12
Silica	0-4.0
Silica HR	0-150
Sulphate	0-200
Sulphide	0-0.5
Sulphite	0-500 (Na <sub>2</sub> SO <sub>3</sub> )
Zinc	0-4.0